

# File Request Form

From

Carl Redmond

Engineer

Date

1/5/98

Phone

6-9698

Attorney

Phone

State

OH

Source Name

Hoover Company

Division or Facility

Division of Maytag

City or County

North Canton

☐ SIP Enforcement Review ☐ CBI ☐ NESHAP ☐ Pollutant

## FOR CASE FILES CHECK APPROPRIATE BOX

☐ Index ☐ 114 Letter/Responses ☐ Inspections by Date ☐ Checklist  
☐ FOV/NOV ☐ 113 Conf Notes ☐ Consent Decree/Order ☐ Complaints/Defendants Response  
☐ Referral ☐ Regulation ☐ Stack Test/Sample/COC For ☒ Misc. CEM Data

## Discovery Section

☐ Interrogatories files separately by compan ☐ Document Produced by USEP  
☐ Deposition of USEP ☐ Documents Produced by Defendant(s)

## CHECK APPROPRIATE BOX: This is a:

☐ Request for Fil Date needed by   
Other   
☐ Request to Open a New File (Be sure to supply info requested above)  
☒ Request to Add Material to Existing Fil

## Special Instructions

This is CEM DATA.

101 East Maple Street  
North Canton, Ohio 44720-2597  
Tel: 330-499-9200

17 September 1997

RECEIVED

SEP 24 1997

AIR ENFORCEMENT BRANCH  
U.S. EPA, REG. 5

Mr. Dave Morehart  
Ohio EPA  
Division of Air Pollution Control  
P.O. Box 1049  
Columbus, Ohio 43216-1049

Dear Mr. Morehart,

To address your letter of September 2, 1997, I have enclosed a sample of the new data recorder chart paper that will be used for the continuous opacity monitoring system from Monitor Labs when the Wickes Boiler B2-2 (Ohio EPA Source 15 76 17 0258 B008) starts up this winter season. This new data recorder chart meets the resolution requirement of 1% divisions that is specified in Table 1-1 of 40 CFR Part 60, Appendix B, Performance Specification 1.

With the installation of the new data recorder chart, the continuous opacity monitoring system from Monitor Labs Model LS-541, Serial Number 0428 meets all of the design and performance specifications of 40 CFR Part 60, Appendix B, Performance Specification 1. So, I feel that the certification report process is complete for this COMS and I should be expecting a certification letter from the agency.

If you have any questions, feel free to call me at 330-499-9200.

Sincerely,



Dal J. Bremer  
Staff Engineer - Environmental



cc: Robert Zahirsky, Canton  
Tammy Hilkins, DAPC  
Patric McCoy, USEPA

